

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)
100.0000 mm	100.0000 mm	0.0000 μm	$\pm 5 \mu\text{m}$	$\pm 5.8\text{E}+00 \mu\text{m}$
105.0000 mm	105.0014 mm	1.4000 μm	$\pm 6 \mu\text{m}$	$\pm 5.8\text{E}+00 \mu\text{m}$
110.0000 mm	110.0020 mm	2.0000 μm	$\pm 6 \mu\text{m}$	$\pm 5.8\text{E}+00 \mu\text{m}$
115.0000 mm	115.0012 mm	1.2000 μm	$\pm 6 \mu\text{m}$	$\pm 5.8\text{E}+00 \mu\text{m}$
120.0000 mm	120.0014 mm	1.4000 μm	$\pm 6 \mu\text{m}$	$\pm 5.8\text{E}+00 \mu\text{m}$
125.0000 mm	125.0012 mm	1.2000 μm	$\pm 6 \mu\text{m}$	$\pm 5.8\text{E}+00 \mu\text{m}$