

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	E.M.P <i>(Tolerance)</i>	Incertidumbre <i>(Uncertainty)</i>	
150.0000 mm	150.0030 mm	3.0000 μm	± FALSE μm	±	5.8E+00 μm
155.0000 mm	155.0014 mm	1.4000 μm	± FALSE μm	±	5.8E+00 μm
160.0000 mm	160.0020 mm	2.0000 μm	± FALSE μm	±	5.8E+00 μm
165.0000 mm	165.0020 mm	2.0000 μm	± FALSE μm	±	5.8E+00 μm
170.0000 mm	170.0012 mm	1.2000 μm	± FALSE μm	±	5.8E+00 μm
175.0000 mm	175.0030 mm	3.0000 μm	± FALSE μm	±	5.8E+00 μm