

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
175.0000 mm	175.0000 mm	0.0000 μm	$\pm 11 \mu\text{m}$	\pm	5.8E+00 μm
180.0000 mm	180.0006 mm	0.6000 μm	$\pm 11 \mu\text{m}$	\pm	5.8E+00 μm
185.0000 mm	185.0014 mm	1.4000 μm	$\pm 11 \mu\text{m}$	\pm	5.8E+00 μm
190.0000 mm	189.9986 mm	-1.4000 μm	$\pm 11 \mu\text{m}$	\pm	5.8E+00 μm
195.0000 mm	194.9986 mm	-1.4000 μm	$\pm 11 \mu\text{m}$	\pm	5.8E+00 μm
200.0000 mm	199.9978 mm	-2.2000 μm	$\pm 11 \mu\text{m}$	\pm	5.8E+00 μm