

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	E.M.P <i>(Tolerance)</i>	Incertidumbre <i>(Uncertainty)</i>	
0.0000 mm	0.0000 mm	0.0000 μm	± 8 μm	±	5.8E+00 μm
90.0000 mm	89.9800 mm	-20.0000 μm	± 9 μm	±	5.8E+00 μm
150.0000 mm	149.9800 mm	-20.0000 μm	± 10 μm	±	5.8E+00 μm
240.0000 mm	239.9800 mm	-20.0000 μm	± 13 μm	±	5.8E+00 μm
300.0000 mm	299.9900 mm	-10.0000 μm	± 15 μm	±	5.8E+00 μm