

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
0.0000 mm	0.0028 mm	2.8000 μm	$\pm 8 \mu\text{m}$	\pm	5.8E+00 μm
5.0000 mm	5.0046 mm	4.6000 μm	$\pm 8 \mu\text{m}$	\pm	5.8E+00 μm
10.0000 mm	10.0070 mm	7.0000 μm	$\pm 8 \mu\text{m}$	\pm	5.8E+00 μm
15.0000 mm	15.0060 mm	6.0000 μm	$\pm 8 \mu\text{m}$	\pm	5.8E+00 μm
20.0000 mm	20.0088 mm	8.8000 μm	$\pm 8 \mu\text{m}$	\pm	5.8E+00 μm
25.0000 mm	25.0088 mm	8.8000 μm	$\pm 8 \mu\text{m}$	\pm	5.8E+00 μm