

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	29.872 °C	29.8 °C	-0.072 °C	± 1.0 °C	± 4.0E-01 °C
35 %HR	34.047 %HR	30.0 %HR	-4.047 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	48.464 %HR	44.0 %HR	-4.464 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	70.540 %HR	66.0 %HR	-4.540 %HR	± 5.0 %HR	± 1.9E+00 %HR