

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
25 °C	25.861 °C	24.8 °C	-1.061 °C	± 2.0 °C	± 4.0E-01 °C
55 %HR	54.824 %HR	55.4 %HR	0.576 %HR	± 5.0 %HR	± 1.9E+00 %HR