

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
22 °C	22.341 °C	20.5 °C	-1.841 °C	± 2.0 °C	± 4.0E-01 °C
33 %HR	37.728 %HR	33.0 %HR	-4.728 %HR	± 5.0 %HR	± 1.9E+00 %HR