

<b>Nominal</b> ( <i>Nominal</i> )	<b>Patrón</b> ( <i>Standard</i> )	<b>UBP</b> ( <i>UUT</i> )	<b>Error</b> ( <i>Error</i> )	<b>Exactitud</b> ( <i>Accuracy</i> )	<b>Incertidumbre</b> ( <i>Uncertainty</i> )
50 °C	49.396 °C	51.2 °C	1.804 °C	± 2.0 °C	± 4.0E-01 °C
60 %HR	56.470 %HR	52.3 %HR	-4.170 %HR	± 5.0 %HR	± 1.9E+00 %HR