

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
53 °C	53.176 °C	53.3 °C	0.124 °C	± 2.0 °C	± 1.2E+00 °C
73 %HR	78.423 %HR	73.6 %HR	-4.823 %HR	± 5.0 %HR	± 1.9E+00 %HR