

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
50 °C	49.696 °C	49.7 °C	0.004 °C	± 2.0 °C	± 4.0E-01 °C
70 %HR	69.865 %HR	65.3 %HR	-4.565 %HR	± 5.0 %HR	± 1.9E+00 %HR