

| Nominal (Nominal) | Patrón (Standard) | UBP (UUT) | Error (Error) | Exactitud (Accuracy) | Incertidumbre (Uncertainty) |
|-------------------|-------------------|-----------|---------------|----------------------|-----------------------------|
| 45 °C             | 44.292 °C         | 44.7 °C   | 0.408 °C      | ± 2.0 °C             | ± 4.0E-01 °C                |
| 67 %HR            | 66.000 %HR        | 64.4 %HR  | -1.600 %HR    | ± 5.0 %HR            | ± 1.9E+00 %HR               |