

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
34 °C	33.833 °C	33.4 °C	-0.433 °C	± 2.0 °C	± 4.0E-01 °C
50 %HR	49.660 %HR	44.8 %HR	-4.860 %HR	± 5.0 %HR	± 1.9E+00 %HR