

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
42 °C	42.226 °C	43.6 °C	1.374 °C	± 2.0 °C	± 4.0E-01 °C
42 %HR	41.296 %HR	37.2 %HR	-4.096 %HR	± 5.0 %HR	± 1.9E+00 %HR