

---

<b>Nominal</b> ( <i>Nominal</i> )	<b>Patrón</b> ( <i>Standard</i> )	<b>UBP</b> ( <i>UUT</i> )	<b>Error</b> ( <i>Error</i> )	<b>Exactitud</b> ( <i>Accuracy</i> )	<b>Incertidumbre</b> ( <i>Uncertainty</i> )
40 °C	40.802 °C	42.7 °C	1.898 °C	± 2.0 °C	± 4.0E-01 °C
44 %HR	43.750 %HR	38.9 %HR	-4.850 %HR	± 5.0 %HR	± 1.9E+00 %HR