

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
50 °C	50.158 °C	50.9 °C	0.742 °C	± 2.0 °C	± 1.2E+00 °C
40 %HR	38.658 %HR	33.9 %HR	-4.758 %HR	± 5.0 %HR	± 1.9E+00 %HR