

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	E.M.P <i>(Tolerance)</i>	Incertidumbre <i>(Uncertainty)</i>	
6.0000 in	6.0000 in	0.0000 in	± 0.0002 in	±	5.8E-05 in
6.0625 in	6.0625 in	0.0000 in	± 0.0002 in	±	5.8E-05 in
6.1000 in	6.1000 in	0.0000 in	± 0.0002 in	±	5.8E-05 in
6.1250 in	6.1251 in	0.0001 in	± 0.0002 in	±	5.8E-05 in
6.2000 in	6.2001 in	0.0001 in	± 0.0002 in	±	5.8E-05 in
6.2500 in	6.2501 in	0.0001 in	± 0.0002 in	±	5.8E-05 in
6.3000 in	6.3001 in	0.0001 in	± 0.0002 in	±	5.8E-05 in
6.5000 in	6.5001 in	0.0001 in	± 0.0002 in	±	5.8E-05 in
7.0000 in	7.0001 in	0.0001 in	± 0.0002 in	±	5.9E-05 in