

Clase de Exactitud (Accuracy Class): Clase III (Class III)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre	Valor UBP	Valor Patrón	Error	EMP	Incertidumbre
m	m	mm	mm	(Uncertainty)	m	m	mm	mm	(Uncertainty)
0 a 0.05	0.04975	0.250		± 4.3E-01 mm					
0 a 0.1	0.09990	0.100		± 4.3E-01 mm					
0 a 0.15	0.15009	-0.090		± 4.3E-01 mm					
0 a 1	0.99860	1.400		± 4.3E-01 mm					
0 a 2	1.99885	1.150		± 4.8E-01 mm					
0 a 3	2.99945	0.550		± 4.8E-01 mm					
0 a 4	3.99984	0.160		± 4.8E-01 mm					
0 a 5	4.99960	0.400		± 4.8E-01 mm					
0 a 6	5.99977	0.230		± 3.6E-01 mm					
0 a 7	6.99964	0.360		± 3.6E-01 mm					
0 a 8	7.99995	0.050		± 3.6E-01 mm					
0 a 9	8.99947	0.530		± 3.6E-01 mm					
0 a 10	9.99958	0.420		± 3.6E-01 mm					
0 a 11	10.99924	0.760		± 3.6E-01 mm					
0 a 12	11.99929	0.710		± 3.6E-01 mm					
0 a 13	12.99985	0.150		± 3.6E-01 mm					
0 a 14	14.00026	-0.260		± 3.6E-01 mm					
0 a 15	15.00009	-0.090		± 3.6E-01 mm					
0 a 16	16.00040	-0.400		± 3.6E-01 mm					
0 a 17	16.99984	0.160		± 3.6E-01 mm					
0 a 18	18.00016	-0.160		± 3.6E-01 mm					
0 a 19	19.00059	-0.590		± 3.6E-01 mm					
0 a 20	20.00086	-0.860		± 3.6E-01 mm					
0 a 21	21.00045	-0.450		± 5.5E-01 mm					
0 a 22	22.00073	-0.730		± 5.5E-01 mm					
0 a 23	23.00005	-0.050		± 5.5E-01 mm					
0 a 24	24.00039	-0.390		± 5.5E-01 mm					
0 a 25	25.00084	-0.840		± 5.5E-01 mm					
0 a 26	26.00065	-0.650		± 5.5E-01 mm					
0 a 27	27.00049	-0.490		± 5.5E-01 mm					
0 a 28	28.00084	-0.840		± 5.5E-01 mm					
0 a 29	29.00058	-0.580		± 5.5E-01 mm					
0 a 30	30.00089	-0.890		± 5.5E-01 mm					