

Clase de Exactitud (Accuracy Class): Clase I (Class I)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre	Valor UBP	Valor Patrón	Error	EMP	Incertidumbre
m	m	mm	mm	(Uncertainty)	m	m	mm	mm	(Uncertainty)
0 a 0.05	0.04945	0.550		± 4.3E-01 mm					
0 a 0.1	0.09962	0.380		± 4.3E-01 mm					
0 a 0.15	0.14998	0.020		± 4.3E-01 mm					
0 a 1	1.00067	-0.670		± 4.3E-01 mm					
0 a 2	1.99915	0.850		± 4.8E-01 mm					
0 a 3	2.99969	0.310		± 4.8E-01 mm					
0 a 4	4.00003	-0.030		± 4.8E-01 mm					
0 a 5	4.99959	0.410		± 4.8E-01 mm					
0 a 6	5.99966	0.340		± 3.6E-01 mm					
0 a 7	6.99958	0.420		± 3.6E-01 mm					
0 a 8	8.00004	-0.040		± 3.6E-01 mm					
0 a 9	8.99971	0.290		± 3.6E-01 mm					
0 a 10	9.99998	0.020		± 3.6E-01 mm					
0 a 11	10.99951	0.490		± 3.6E-01 mm					
0 a 12	11.99967	0.330		± 3.6E-01 mm					
0 a 13	12.99970	0.300		± 3.6E-01 mm					
0 a 14	13.99993	0.070		± 3.6E-01 mm					
0 a 15	14.99996	0.040		± 3.6E-01 mm					
0 a 16	16.00009	-0.090		± 3.6E-01 mm					
0 a 17	16.99941	0.590		± 3.6E-01 mm					
0 a 18	17.99957	0.430		± 3.6E-01 mm					
0 a 19	18.99946	0.540		± 3.6E-01 mm					
0 a 20	19.99951	0.490		± 3.6E-01 mm					
0 a 21	21.00004	-0.040		± 5.5E-01 mm					
0 a 22	22.00034	-0.340		± 5.5E-01 mm					
0 a 23	23.00025	-0.250		± 5.5E-01 mm					
0 a 24	24.00048	-0.480		± 5.5E-01 mm					
0 a 25	25.00091	-0.910		± 5.5E-01 mm					
0 a 26	26.00031	-0.310		± 5.5E-01 mm					
0 a 27	27.00019	-0.190		± 5.5E-01 mm					
0 a 28	28.00044	-0.440		± 5.5E-01 mm					
0 a 29	29.00021	-0.210		± 5.5E-01 mm					
0 a 30	30.00061	-0.610		± 5.5E-01 mm					