

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	30.373 °C	30.8 °C	0.427 °C	± 1.0 °C	± 4.0E-01 °C
35 %HR	35.181 %HR	29.1 %HR	-6.081 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	50.693 %HR	42.8 %HR	-7.893 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	70.315 %HR	63.6 %HR	-6.715 %HR	± 5.0 %HR	± 1.9E+00 %HR