

Clase de Exactitud (Accuracy Class): Clase I (Class I)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre	Valor UBP	Valor Patrón	Error	EMP	Incertidumbre
ft	ft	in	in	(Uncertainty)	ft	ft	in	in	(Uncertainty)
0 a 5	4.99905	0.011	± 0.10	± 1.4E-03 in					
0 a 10	9.99955	0.005	± 2.2	± 1.4E-03 in					
0 a 15	15.00136	-0.016	± 3.2	± 1.4E-03 in					
0 a 20	20.00088	-0.011	± 4.2	± 1.4E-03 in					
0 a 25	25.00149	-0.018	± 5.2	± 1.4E-03 in					
0 a 30	30.00195	-0.023	± 6.2	± 1.4E-03 in					
0 a 35	35.00092	-0.011	± 7.2	± 1.4E-03 in					
0 a 40	40.00135	-0.016	± 8.2	± 1.4E-03 in					
0 a 45	45.00269	-0.032	± 9.2	± 1.4E-03 in					
0 a 50	50.00127	-0.015	± 10.2	± 1.4E-03 in					