

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
20 °C	19.814 °C	20.3 °C	0.486 °C	± 1.0 °C	± 4.0E-01 °C
25 °C	24.798 °C	25.2 °C	0.402 °C	± 1.0 °C	± 4.0E-01 °C
30 °C	29.872 °C	30.3 °C	0.428 °C	± 1.0 °C	± 4.0E-01 °C
35 °C	34.896 °C	35.3 °C	0.404 °C	± 1.0 °C	± 4.0E-01 °C
40 °C	39.919 °C	40.3 °C	0.381 °C	± 1.0 °C	± 4.0E-01 °C
60 %HR	61.848 %HR	60.1 %HR	-1.748 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.366 %HR	68.4 %HR	-1.966 %HR	± 3.0 %HR	± 1.9E+00 %HR
80 %HR	80.713 %HR	77.9 %HR	-2.813 %HR	± 3.0 %HR	± 1.9E+00 %HR
90 %HR	89.732 %HR	86.7 %HR	-3.032 %HR	± 3.0 %HR	± 1.9E+00 %HR