

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
22 °C	22.191 °C	22.4 °C	0.209 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.450 °C	25.6 °C	0.150 °C	± 0.5 °C	± 4.0E-01 °C
31 °C	31.366 °C	31.7 °C	0.334 °C	± 0.5 °C	± 4.0E-01 °C
48 %HR	48.331 %HR	48.9 %HR	0.569 %HR	± 3.0 %HR	± 1.9E+00 %HR
57 %HR	57.666 %HR	58.3 %HR	0.634 %HR	± 3.0 %HR	± 1.9E+00 %HR
67 %HR	67.073 %HR	67.2 %HR	0.127 %HR	± 3.0 %HR	± 1.9E+00 %HR