

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
22 °C	22.191 °C	22.2 °C	0.009 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.450 °C	25.5 °C	0.050 °C	± 0.5 °C	± 4.0E-01 °C
31 °C	31.366 °C	31.6 °C	0.234 °C	± 0.5 °C	± 4.0E-01 °C
48 %HR	48.331 %HR	50.2 %HR	1.869 %HR	± 3.0 %HR	± 1.9E+00 %HR
57 %HR	57.666 %HR	58.2 %HR	0.534 %HR	± 3.0 %HR	± 1.9E+00 %HR
67 %HR	67.073 %HR	66.3 %HR	-0.773 %HR	± 3.0 %HR	± 1.9E+00 %HR