

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
22 °C	22.191 °C	22.2 °C	0.009 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.450 °C	25.5 °C	0.050 °C	± 0.5 °C	± 4.0E-01 °C
31 °C	31.366 °C	31.6 °C	0.234 °C	± 0.5 °C	± 4.0E-01 °C
48 %HR	48.597 %HR	48.3 %HR	-0.297 %HR	± 3.0 %HR	± 1.9E+00 %HR
57 %HR	57.666 %HR	58.3 %HR	0.634 %HR	± 3.0 %HR	± 1.9E+00 %HR
67 %HR	67.073 %HR	67.3 %HR	0.227 %HR	± 3.0 %HR	± 1.9E+00 %HR