

<b>Nominal (Nominal)</b>	<b>Patrón (Standard)</b>	<b>UBP (UUT)</b>	<b>Error (Error)</b>	<b>Exactitud (Accuracy)</b>	<b>Incertidumbre (Uncertainty)</b>
22 °C	22.191 °C	22.4 °C	0.209 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.450 °C	25.6 °C	0.150 °C	± 0.5 °C	± 4.0E-01 °C
31 °C	31.366 °C	31.6 °C	0.234 °C	± 0.5 °C	± 4.0E-01 °C
48 %HR	48.331 %HR	48.2 %HR	-0.131 %HR	± 3.0 %HR	± 1.9E+00 %HR
57 %HR	57.666 %HR	57.5 %HR	-0.166 %HR	± 3.0 %HR	± 1.9E+00 %HR
67 %HR	67.073 %HR	66.2 %HR	-0.873 %HR	± 3.0 %HR	± 1.9E+00 %HR