

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
2 °C	2.085 °C	2.3 °C	0.215 °C	± 0.5 °C	± 4.0E-01 °C
5 °C	5.394 °C	5.6 °C	0.206 °C	± 0.5 °C	± 4.0E-01 °C
8 °C	8.302 °C	8.4 °C	0.098 °C	± 0.5 °C	± 4.0E-01 °C