

<b>Nominal (Nominal)</b>	<b>Patrón (Standard)</b>	<b>UBP (UUT)</b>	<b>Error (Error)</b>	<b>Exactitud (Accuracy)</b>	<b>Incertidumbre (Uncertainty)</b>
22 °C	22.261 °C	22.3 °C	0.039 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.049 °C	24.8 °C	-0.249 °C	± 0.5 °C	± 4.0E-01 °C
31 °C	31.286 °C	31.2 °C	-0.086 °C	± 0.5 °C	± 4.0E-01 °C
48 %HR	48.331 %HR	48.8 %HR	0.469 %HR	± 3.0 %HR	± 1.9E+00 %HR
57 %HR	57.257 %HR	57.8 %HR	0.543 %HR	± 3.0 %HR	± 1.9E+00 %HR
67 %HR	67.687 %HR	68.2 %HR	0.513 %HR	± 3.0 %HR	± 1.9E+00 %HR