

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
22 °C	22.261 °C	22.2 °C	-0.061 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.049 °C	24.9 °C	-0.149 °C	± 0.5 °C	± 4.0E-01 °C
31 °C	31.286 °C	31.3 °C	0.014 °C	± 0.5 °C	± 4.0E-01 °C
48 %HR	48.331 %HR	48.5 %HR	0.169 %HR	± 3.0 %HR	± 1.9E+00 %HR
57 %HR	57.257 %HR	57.7 %HR	0.443 %HR	± 3.0 %HR	± 1.9E+00 %HR
67 %HR	67.687 %HR	67.4 %HR	-0.287 %HR	± 3.0 %HR	± 1.9E+00 %HR