

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
22 °C	22.261 °C	22.1 °C	-0.161 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.049 °C	24.9 °C	-0.149 °C	± 0.5 °C	± 4.0E-01 °C
31 °C	31.286 °C	31.2 °C	-0.086 °C	± 0.5 °C	± 4.0E-01 °C
48 %HR	48.331 %HR	48.4 %HR	0.069 %HR	± 3.0 %HR	± 1.9E+00 %HR
57 %HR	57.257 %HR	57.7 %HR	0.443 %HR	± 3.0 %HR	± 1.9E+00 %HR
67 %HR	67.687 %HR	67.6 %HR	-0.087 %HR	± 3.0 %HR	± 1.9E+00 %HR