

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
22 °C	22.261 °C	22.5 °C	0.239 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.049 °C	25.0 °C	-0.049 °C	± 0.5 °C	± 4.0E-01 °C
31 °C	31.286 °C	31.4 °C	0.114 °C	± 0.5 °C	± 4.0E-01 °C
48 %HR	48.331 %HR	48.6 %HR	0.269 %HR	± 3.0 %HR	± 1.9E+00 %HR
57 %HR	57.257 %HR	56.8 %HR	-0.457 %HR	± 3.0 %HR	± 1.9E+00 %HR
67 %HR	67.687 %HR	67.5 %HR	-0.187 %HR	± 3.0 %HR	± 1.9E+00 %HR