

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	31.246 °C	31.7 °C	0.454 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	32.860 %HR	35.2 %HR	2.340 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	51.184 %HR	52.1 %HR	0.916 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	69.977 %HR	70.6 %HR	0.663 %HR	± 3.0 %HR	± 1.9E+00 %HR