

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	E.M.P <i>(Tolerance)</i>	Incertidumbre <i>(Uncertainty)</i>	
23.30 μm	22 μm	-1.30 μm	± 5 μm	±	5.8E-01 μm
51.60 μm	51 μm	-0.60 μm	± 5 μm	±	5.8E-01 μm
124.60 μm	125 μm	0.40 μm	± 6 μm	±	5.8E-01 μm
250.30 μm	249 μm	-1.30 μm	± 13 μm	±	5.8E-01 μm
486.00 μm	487 μm	1.00 μm	± 24 μm	±	5.8E-01 μm