

| Patrón (Standard) | UBP (UUT) | Error (Error) | E.M.P (Tolerance) | Incertidumbre (Uncertainty) |            |
|-------------------|-----------|---------------|-------------------|-----------------------------|------------|
| 23.3 μm           | 23.4 μm   | 0.1 μm        | ± 2.5 μm          | ±                           | 5.8E-02 μm |
| 51.6 μm           | 51.4 μm   | -0.2 μm       | ± 2.5 μm          | ±                           | 5.8E-02 μm |
| 124.6 μm          | 123.0 μm  | -1.6 μm       | ± 3.7 μm          | ±                           | 6.1E-02 μm |
| 250.3 μm          | 251 μm    | 0.7 μm        | ± 7.5 μm          | ±                           | 6.9E-02 μm |
| 486.0 μm          | 486 μm    | 0.0 μm        | ± 14.6 μm         | ±                           | 9.4E-02 μm |
| 963.0 μm          | 962 μm    | -1.0 μm       | ± 29 μm           | ±                           | 1.6E-01 μm |