

Clase de Exactitud (Accuracy Class): Clase III (Class III)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre	Valor UBP	Valor Patrón	Error	EMP	Incertidumbre
m	m	mm	mm	(Uncertainty)	m	m	mm	mm	(Uncertainty)
0 a 0.6	0.600	0.27	± 1.6	± 4.3E-01 mm					
0 a 0.7	0.700	0.27	± 1.7	± 4.3E-01 mm					
0 a 0.8	0.800	0.27	± 1.8	± 4.3E-01 mm					
0 a 0.9	0.900	0.28	± 1.9	± 4.3E-01 mm					
0 a 1	1.000	0.28	± 2	± 4.3E-01 mm					
0 a 1.1	1.100	0.28	± 2	± 4.8E-01 mm					
0 a 1.2	1.200	0.29	± 2.1	± 4.8E-01 mm					
0 a 1.3	1.300	0.29	± 2.2	± 4.8E-01 mm					
0 a 1.4	1.400	0.30	± 2.3	± 4.8E-01 mm					
0 a 1.5	1.499	1.30	± 2.4	± 4.8E-01 mm					
0 a 1.6	1.599	1.31	± 2.4	± 4.8E-01 mm					
0 a 1.7	1.699	1.31	± 2.5	± 4.8E-01 mm					
0 a 1.8	1.799	1.32	± 2.6	± 4.8E-01 mm					
0 a 1.9	1.899	1.00	± 2.7	± 4.8E-01 mm					
0 a 2	1.999	1.00	± 2.8	± 4.8E-01 mm					