

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
15 °C	15.121 °C	15.4 °C	0.279 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.363 °C	30.2 °C	-0.163 °C	± 3.0 °C	± 4.0E-01 °C
50 %HR	50.887 %HR	52.6 %HR	1.713 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.315 %HR	72.6 %HR	2.285 %HR	± 3.0 %HR	± 1.9E+00 %HR