

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
2 °C	2.246 °C	2.4 °C	0.154 °C	± 0.5 °C	± 4.0E-01 °C
5 °C	5.154 °C	5.0 °C	-0.154 °C	± 0.5 °C	± 4.0E-01 °C
8 °C	8.262 °C	8.1 °C	-0.162 °C	± 0.5 °C	± 4.0E-01 °C