

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
18 °C	18.210 °C	18.1 °C	-0.130 °C	± 0.2 °C	± 4.0E-01 °C
20 °C	20.085 °C	20.2 °C	0.135 °C	± 0.2 °C	± 4.0E-01 °C
25 °C	25.610 °C	25.5 °C	-0.130 °C	± 0.2 °C	± 4.0E-01 °C
35 °C	35.076 °C	35.2 °C	0.134 °C	± 0.2 °C	± 4.0E-01 °C
40 %HR	40.509 %HR	37.2 %HR	-3.309 %HR	± 3.5 %HR	± 1.9E+00 %HR
75 %HR	75.304 %HR	75.1 %HR	-0.254 %HR	± 3.5 %HR	± 1.9E+00 %HR
80 %HR	80.059 %HR	80.2 %HR	0.141 %HR	± 3.5 %HR	± 1.9E+00 %HR