

Clase de Exactitud (Accuracy Class): Clase I (Class I)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre	Valor UBP	Valor Patrón	Error	EMP	Incertidumbre
m	m	mm	mm	(Uncertainty)	m	m	mm	mm	(Uncertainty)
0 a 0.03	0.02994	0.060	± 0.2	± 4.3E-01 mm					
0 a 0.06	0.05998	0.020	± 0.2	± 4.3E-01 mm					
0 a 0.09	0.08997	0.030	± 0.2	± 4.3E-01 mm					
0 a 0.12	0.12000	0.000	± 0.2	± 4.3E-01 mm					
0 a 0.15	0.15005	-0.050	± 0.2	± 4.3E-01 mm					
0 a 0.18	0.18008	-0.080	± 0.2	± 4.3E-01 mm					
0 a 0.21	0.21007	-0.070	± 0.2	± 4.3E-01 mm					
0 a 0.24	0.24009	-0.090	± 0.2	± 4.3E-01 mm					
0 a 0.27	0.27010	-0.100	± 0.2	± 4.3E-01 mm					
0 a 0.3	0.30012	-0.120	± 0.2	± 4.3E-01 mm					
