
Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
52.65 μm	52.68 μm	0.03 μm	$\pm 2.5 \mu\text{m}$	\pm	5.8E-01 μm
155.23 μm	155.0 μm	-0.23 μm	$\pm 2.5 \mu\text{m}$	\pm	5.8E-01 μm
257.12 μm	257.2 μm	0.08 μm	$\pm 2.5 \mu\text{m}$	\pm	5.8E-01 μm
511.72 μm	511.0 μm	-0.72 μm	$\pm 2.5 \mu\text{m}$	\pm	5.8E-01 μm