

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.293 °C	30.1 °C	-0.193 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	30.570 %HR	30.0 %HR	-0.570 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	50.887 %HR	54.0 %HR	3.113 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	70.315 %HR	74.0 %HR	3.685 %HR	± 5.0 %HR	± 1.9E+00 %HR