

<b>Nominal (Nominal)</b>	<b>Patrón (Standard)</b>	<b>UBP (UUT)</b>	<b>Error (Error)</b>	<b>Exactitud (Accuracy)</b>	<b>Incertidumbre (Uncertainty)</b>
30 °C	30.293 °C	30.0 °C	-0.293 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	30.570 %HR	33.0 %HR	2.430 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	50.887 %HR	55.0 %HR	4.113 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	70.315 %HR	75.0 %HR	4.685 %HR	± 5.0 %HR	± 1.9E+00 %HR