

| Nominal ( <i>Nominal</i> ) | Patrón ( <i>Standard</i> ) | UBP ( <i>UUT</i> ) | Error ( <i>Error</i> ) | Exactitud ( <i>Accuracy</i> ) | Incertidumbre ( <i>Uncertainty</i> ) |
|----------------------------|----------------------------|--------------------|------------------------|-------------------------------|--------------------------------------|
| 10 °C                      | 10.148 °C                  | 9.8 °C             | -0.348 °C              | ± 0.5 °C                      | ± 4.0E-01 °C                         |
| 20 °C                      | 20.356 °C                  | 19.9 °C            | -0.456 °C              | ± 0.5 °C                      | ± 4.0E-01 °C                         |
| 30 °C                      | 30.313 °C                  | 29.9 °C            | -0.413 °C              | ± 0.5 °C                      | ± 4.0E-01 °C                         |
| 40 °C                      | 40.040 °C                  | 39.9 °C            | -0.140 °C              | ± 0.5 °C                      | ± 4.0E-01 °C                         |