

<b>Nominal</b> ( <i>Nominal</i> )	<b>Patrón</b> ( <i>Standard</i> )	<b>UBP</b> ( <i>UUT</i> )	<b>Error</b> ( <i>Error</i> )	<b>Exactitud</b> ( <i>Accuracy</i> )	<b>Incertidumbre</b> ( <i>Uncertainty</i> )
10 °C	10.148 °C	10.1 °C	-0.048 °C	± 0.5 °C	± 4.0E-01 °C
20 °C	20.356 °C	20.1 °C	-0.256 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.313 °C	30.1 °C	-0.213 °C	± 0.5 °C	± 4.0E-01 °C
40 °C	40.040 °C	39.9 °C	-0.140 °C	± 0.5 °C	± 4.0E-01 °C