

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
2 °C	2.587 °C	5.2 °C	2.613 °C	± 1.0 °C	± 4.0E-01 °C
8 °C	8.353 °C	10.2 °C	1.847 °C	± 1.0 °C	± 4.0E-01 °C
30 °C	30.062 °C	29.5 °C	-0.562 °C	± 1.0 °C	± 4.0E-01 °C
50 %HR	50.253 %HR	55.0 %HR	4.747 %HR	± 5.0 %HR	± 1.9E+00 %HR