

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.544 °C	30.5 °C	-0.044 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.529 %HR	32.5 %HR	1.971 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.366 %HR	52.7 %HR	2.334 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.243 %HR	73.2 %HR	2.957 %HR	± 3.0 %HR	± 1.9E+00 %HR