

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	30.062 °C	30.3 °C	0.238 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	29.179 %HR	38.5 %HR	9.321 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	49.681 %HR	59.2 %HR	9.519 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	69.425 %HR	70.4 %HR	0.975 %HR	± 3.0 %HR	± 1.9E+00 %HR