

<b>Nominal</b> ( <i>Nominal</i> )	<b>Patrón</b> ( <i>Standard</i> )	<b>UBP</b> ( <i>UUT</i> )	<b>Error</b> ( <i>Error</i> )	<b>Exactitud</b> ( <i>Accuracy</i> )	<b>Incertidumbre</b> ( <i>Uncertainty</i> )
15 °C	15.151 °C	15.4 °C	0.249 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	29.812 °C	29.6 °C	-0.212 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	31.460 %HR	32.7 %HR	1.240 %HR	± 2.0 %HR	± 1.9E+00 %HR
70 %HR	71.010 %HR	72.9 %HR	1.890 %HR	± 2.0 %HR	± 1.9E+00 %HR