

Certificado No. (Certificate #):53118Equipo (Instrument):INDICADOR DE DIAL

Fecha de Calibración (Calibration Date):2025-01-21Marca (Brand):Mitutoyo

Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)	
1.0000 mm	1.0100 mm	10.0000 µm	± 13 µm	±	5.8E+00 µm
1.0500 mm	1.0600 mm	10.0000 µm	± 13 µm	±	5.8E+00 µm
1.1000 mm	1.1100 mm	10.0000 µm	± 13 µm	±	5.8E+00 µm
1.5000 mm	1.5100 mm	10.0000 µm	± 13 µm	±	5.8E+00 µm
2.0000 mm	2.0100 mm	10.0000 µm	± 13 µm	±	5.8E+00 µm
2.5000 mm	2.5100 mm	10.0000 µm	± 13 µm	±	5.8E+00 µm
4.0000 mm	4.0000 mm	0.0000 µm	± 13 µm	±	5.8E+00 µm
4.5000 mm	4.5000 mm	0.0000 µm	± 13 µm	±	5.8E+00 µm
5.0000 mm	5.0000 mm	0.0000 µm	± 13 µm	±	5.8E+00 µm