

Certificado No. (*Certificate #*): 53118  
Fecha de Calibración (*Calibration Date*): 2025-01-21

Equipo  
(*Instrument*): INDICADOR DE DIAL  
Marca (*Brand*): Mitutoyo

<b>Patrón</b> ( <i>Standard</i> )	<b>UBP</b> ( <i>UUT</i> )	<b>Error</b> ( <i>Error</i> )	<b>Exactitud</b> ( <i>Accuracy</i> )	<b>Incertidumbre</b> ( <i>Uncertainty</i> )
1.0000 mm	1.0100 mm	10.0000 $\mu\text{m}$	$\pm 13 \mu\text{m}$	$\pm 5.8\text{E}+00 \mu\text{m}$
1.0500 mm	1.0600 mm	10.0000 $\mu\text{m}$	$\pm 13 \mu\text{m}$	$\pm 5.8\text{E}+00 \mu\text{m}$
1.1000 mm	1.1100 mm	10.0000 $\mu\text{m}$	$\pm 13 \mu\text{m}$	$\pm 5.8\text{E}+00 \mu\text{m}$
1.5000 mm	1.5100 mm	10.0000 $\mu\text{m}$	$\pm 13 \mu\text{m}$	$\pm 5.8\text{E}+00 \mu\text{m}$
2.0000 mm	2.0100 mm	10.0000 $\mu\text{m}$	$\pm 13 \mu\text{m}$	$\pm 5.8\text{E}+00 \mu\text{m}$
2.5000 mm	2.5100 mm	10.0000 $\mu\text{m}$	$\pm 13 \mu\text{m}$	$\pm 5.8\text{E}+00 \mu\text{m}$
4.0000 mm	4.0000 mm	0.0000 $\mu\text{m}$	$\pm 13 \mu\text{m}$	$\pm 5.8\text{E}+00 \mu\text{m}$
4.5000 mm	4.5000 mm	0.0000 $\mu\text{m}$	$\pm 13 \mu\text{m}$	$\pm 5.8\text{E}+00 \mu\text{m}$
5.0000 mm	5.0000 mm	0.0000 $\mu\text{m}$	$\pm 13 \mu\text{m}$	$\pm 5.8\text{E}+00 \mu\text{m}$